


## Test Certificate

Part 1 Liter Vessel  
 Part Number 14143-01  
 Serial Number 105051 - 105282



		USP / Ph.Eur. Tolerances	
<b>Dimensions</b>	<b>Vessel</b>	Inside Ø a	102.9 to 105.1 mm
		Height b	200 mm
With hemispherical bottom			
<b>Material</b>	Borosilicate glass		

The accuracy of the above mentioned part dimensions have been examined. All relevant measurements are within tolerance and the factory acceptance test has passed. The dimensions are in conformance with USP and Ph. Eur. Pharmacopoeia requirements (at date of approval, see below).

Aesch, 2026-01-06

Buner Cornelius  
 SOTAX Quality Management  
 SOTAX AG | Nordring 1 | 4147 Aesch | Switzerland